

CENTRAL QUEENSLAND SEISMOLOGY RESEARCH GROUP

# CQSRG Seismological Report 2014

---

Edition 2015-09-21

**Compiled by: Michael Turnbull BAppSc(Distinction) QUT MAppSc CQU, July 2015.**

## Table of Contents

Copyright Statement.....	i
Edition Audit Trace .....	ii
Introduction .....	1
FS03 Uptime during 2014 .....	1
Earthquake Events Detected during 2014 .....	2
Public Seismic Network (PSN) .....	3
Appendix A – Details of FS03 .....	4
Appendix B – CQSRG Method of Earthquake Location .....	5
Appendix C – CQSRG Method of Magnitude Quantification .....	6
Calibration of FS03 Seismometer for Earthquake Magnitude Determination.....	6
Introduction .....	6
Background Information .....	6
Method .....	7
Important Note Concerning Accuracy and Precision .....	8
Example Usage.....	9
Student Resources .....	10

## Copyright Statement

The format of this document is Copyright © 2015 Central Queensland Seismology Research Group (CQSRG). CQSRG reserves all rights to the document format.

The contents of this document are released to the public domain. The information contained in this document may be used freely; however, CQSRG requests that the normal academic practice of citing this document when used as an information source be respected.

This document may be cited as follows.

(CQSRG 2015), *CQSRG Seismological Report 2014*, Edition 20150921, Central Queensland Seismology Research Group, Compiler: Michael Turnbull, Published September 2015 at <http://www.cqsrg.org>.

Michael Turnbull may be contacted by email at: [Mike.Turnbull@cqsrg.org](mailto:Mike.Turnbull@cqsrg.org)

## Edition Audit Trace

<b>Edition Number</b>	<b>Date Published</b>	<b>Editorial Changes Made</b>	<b>Editor</b>
	July 2015	Original edition	MLT
20150921	21 September 2015	Minor typographic errors corrected. Bolding added to Table 1. Contact email address changed. Table of Contents and Recommended citation altered.	MLT

## Introduction

This report details earthquakes detected and located by the Central Queensland Seismology Research Group (CQSRG) in 2014. Detailed summaries of earthquakes that occurred or were felt in Central Queensland are provided. The date and time of earthquakes noted in this report are provided in Universal Coordinated Time (UTC).

CQSRG was established in 2002, under the auspices of the Faculty of Informatics and Communication of Central Queensland University (CQU), with Michael Turnbull (Lecturer, and later Adjunct Research Fellow) and Kevin McCue (Visiting Professor, and later Adjunct Professor) as the designated researchers. This affiliation with CQU continued until 2013, when, due to a divergence in academic focus of CQU and CQSRG, the researchers allowed their Adjunct appointments to lapse. Subsequently CQSRG continued independent of CQU, with the same two people conducting the research.

During the 2014 calendar year CQSRG operated one seismic monitoring station, designated FS03. Details of this station, including location and equipment, are provided in Appendix A. This report only contains information relating to earthquakes detected by the FS03 seismic monitoring station.

CQSRG locates and quantifies earthquakes using the methods detailed in Appendices B and C.




Figure 1: Percentage Uptime/Downtime of FS03.

### FS03 Uptime during 2014

During 2014 the FS03 Seismic Monitoring Station was actively on line and monitoring for events 62% of the overall year. This is depicted in Figure 1.

The main cause of downtime was the memory filling up and preventing further storage of triggered events. Downloading of triggered event files, and subsequent clearing of the FS03 storage memory has been conducted manually since it was installed in 2004.

Consequently this operational procedure is dependent on the availability of trained personnel. Currently there is only one such person. The FS03 trigger settings are such that the available 6 MiB<sup>1</sup> of storage memory is exhausted in approximately 2 weeks, and takes about an hour to download.

Some downtime was caused by power failures.

---

<sup>1</sup> The incorrect use of standard SI decade multiplier prefixes as binary multipliers has caused a great deal of confusion over a long period. For example, does a 6 megabyte file contain 6 000 000 ( $6 \times 10^6$ ) or 6 291 456 ( $6 \times 2^{20}$ ) bytes? This confusion has now been resolved - at least for those willing to adopt an international standard. In December 1998, the International Electrotechnical Commission (IEC) approved standard names and prefixes for binary multipliers. NIST (National Institute of Standards and Technology) has published the full set of standard names and prefixes for binary multipliers at <http://physics.nist.gov/cuu/Units/binary.html>.

Downtime periods ranged from as short as 25 minutes during data download sessions, to as long as 60 days, during a prolonged absence of the operator.

There are plans in place to automate the data download and memory clearance.


## Earthquake Events Detected during 2014

During 2014 only six earthquake events were detected by CQSRG, and locations were attempted for all events. Details of these events are provided in Table 1.

**Table 1: Earthquake Events Detected by CQSRG During 2014.**

Date (UTC)	Time (UTC)	Longitude <sup>2</sup>	Latitude	Depth	Depth Status	Magnitude (ML)	Event Type	Place	Comment
2014-06-14	14:19:07.28	151.655	-25.108	10.0 km	Constrained to local norm	3.0	Earthquake Mainshock	Mt Perry	8 km N Mt Perry. Reviewed 2015-07-09.
2014-08-22	08:34	150.72	-25.71	10.0 km	Constrained to local norm	1.9	Earthquake Foreshock	Nerangi	Approximately 75 km NE Taroom, 60 km SE Eidsvold. 7 km SW Nerangi station. Not heard at Nerangi.
2014-08-22	08:35	150.72	-25.71	10.0 km	Constrained to local norm	2.7	Earthquake Foreshock	Nerangi	Approximately 75 km NE Taroom, 60 km SE Eidsvold. 7 km SW Nerangi station. Felt & heard at Nerangi.
2014-08-22	08:38	150.72	-25.71	10.0 km	Constrained to local norm	2.8	Earthquake Mainshock	Nerangi	Approximately 75 km NE Taroom, 60 km SE Eidsvold. 7 km SW Nerangi station. Felt & heard at Nerangi.
2014-08-26	11:01:40.81			10.0 km	Constrained to local norm	2.4	Earthquake Mainshock		Ambiguous location (-23.832, 152.570) or (-25.545, 153.409). Reviewed 2015-07-09.
2014-08-27	21:02	150.72	-25.71	10.0 km	Constrained to local norm	2.7	Earthquake Aftershock	Nerangi	Approximately 75 km NE Taroom, 60 km SE Eidsvold. 7 km SW Nerangi station. Felt & heard at Nerangi.

Figure 2 provides a graphical representation of the frequency of magnitude spread, while Figure 3 is a map of the event locations. Note that the location for the 2014-08-26 event is ambiguous between one of two possible locations. Further data is required to disambiguate the location.



**Figure 2: 2014 Earthquake Magnitude Frequency.**

<sup>2</sup> The precision of the latitude and longitude values used in Table 1 are as provided by the EQLOCL algorithm and may not be a true indication of the actual locational error range.

The red circle in Figure 3 indicates that, during 2014, the most distant earthquake event detected by CQSRG was ML 2.4, approximately 160 km from FS03.




Figure 3: 2014 Earthquake Locations.

## Public Seismic Network (PSN)

Since 2011-08-05 CQSRG has hosted a PSN seismograph station, known to the Australian PSN community as the Gin Gin or the Horse Camp station. Vic Dent and Mike Turnbull originally installed the station with a rudimentary setup consisting of a 3D geophone attached to a PSN A/D board, in a vacant brick shed on Mike Turnbull's property at Horse Camp, 16 km SW of Gin Gin. Mike provided a desktop computer onto which the PSN software was installed. Since then the station has been regularly uploading GIF pictures of the daily seismogram traces to the Regional Seismic Users web site at <http://www.rsuw.daleh.id.au/>, operated by Dale Hardy. The station also uploads continuous data to the Regional Seismic Network (RSN), operated by the Australian Centre for Geomechanics (AGC) (Information at <http://www.acg.uwa.edu.au/>).

In 2013 the geophone was replaced with a Sprengnether S6000 seismometer, and the PSN A/D board was housed in a respectable electronics housing, along with custom made adaptor electronics to accommodate the sensor and GPS interface.

Since the PSN station is located only 300 m from FS03, data from the PSN station is not used in locating events detected by CQSRG. This may change if negotiations with the Gin Gin State High School result in the PSN station being transferred to their campus in Gin Gin.

## Appendix A – Details of FS03

### LOCATION

Latitude -25.1068, Longitude 151.8667, Height above sea level 180 m. Approximately 16 km SW Gin Gin, Queensland, Australia.

### SITE AND SAMPLING

Sampling of ground velocity at 100 sample/sec, full scale 4194304 counts

Ch Type	Serial	Name	Direction	Gain	Filters
1 L43D	#1482	East	90 deg true	0.00	DC 50.0
2 L43D	#1482	North	0 deg true	0.00	DC 50.0
3 L43D	#1482	Up	Positive up	0.00	DC 50.0

### DATA LOGGER

Kelunji Classic #153, GURIA V4.16A Operating System.

### TIME SYNC

Sync every day at 1400 UCT, using GPS. Wait for up to 80 seconds

Wait up to 120 seconds for a position

Auto-correct clock after sync

### TRIGGER SETTINGS

STA/LTA Channel 3, filter 1.00 to 7.50 Hz

Time const 0.20, 2.0, 20.0, 200.0 seconds

Ratios Fast 3.50, slow 1.75, squelch 5, 15 days

Length 100 to 200 secs, 80.00 sec pre-trigger, 1.10 cutoff.

## Appendix B – CQSRG Method of Earthquake Location

In general, CQSRG only catalogues earthquake events that are detected by its seismic monitoring station(s). However, in the event of significant local events that, for reasons of station downtime, are not recorded by CQSRG stations, locations are conducted by obtaining data from other agencies.

The general process for earthquake event location at CQSRG is as follows.

1. Identify local earthquake events from visual inspection of FS03 seismograms.
2. Download extra seismograms from other agencies; typically, University of Queensland, Geoscience Australia, and the Australian National University (ANU) Australian Seismometers in Schools (AuSIS) project.
3. Send email requests to other agencies; typically, the Seismology Research Centre (SRC), and the South East Queensland Water Company (SeqWater).
4. Collect all available seismogram records and pick P and S phase arrival times using EqWave (SRC sourced software).
5. Enter the picked P and S times into EQLOCL (SRC sourced software).
6. Use the location calculated by EQLOCL.

In the not so rare cases where the only record available is that from FS03, an attempt is made to locate the event using first motion polarity and near field trigonometry. This can only be done when the first motions are sufficiently impulsive to give an unambiguous indication of the arrival azimuth.

In cases where only two records are available (invariably FS03 and EIDS), and the S-P derived radial distance circles meet, but do not over extend, the touch point is used as a seed to the EQLOCL algorithm.

In cases where only two records are available (invariably FS03 and EIDS), and the S-P derived radial distance circles over extend, but the first motions are sufficiently impulsive to derive an unambiguous azimuth, the radial touch point indicated by the azimuth direction is used as a seed to the EQLOCL algorithm.

In cases where only two records are available (invariably FS03 and EIDS), and the S-P derived radial distance circles over extend, but the first motions are insufficiently impulsive to derive an unambiguous azimuth, the locations of both the radial touch points are used as seeds to the EQLOCL algorithm, and the resulting ambiguous locations are noted in the catalogue entry comments.

In cases where the only information that can be gleaned is the radial distance from FS03, that distance may be noted in the catalogue listing comments.



## Appendix C – CQSRG Method of Magnitude Quantification

### Calibration of FS03 Seismometer for Earthquake Magnitude Determination.

Mike Turnbull, 7 November, 2012.

#### Introduction

FS03 is the designation of a seismic monitoring station operated by the Central Queensland Seismology Research Group (CQSRG). It is located about 16 km south-west of Gin Gin.

When the FS03 station was first installed it had a Sprengnether S6000 seismometer attached to a data logger manufactured by the Seismology Research Centre (SRC). The characteristics of this sensor and the amplification factors of the data logger section of the seismograph were used as input to the SRC software used to locate and quantify earthquakes recorded on the seismograph. When the S6000 sensor failed it was replaced with a Mark Products L43D seismometer sensor. By comparison of the calibration waveform amplitudes of the S6000 against the L43D, a correction factor of 1.7 was calculated and used to adjust the amplitude value input to the SRC software to determine earthquake magnitudes using the new sensor – and this provided a temporary solution.

In order for the SRC software to be able to calculate an earthquake magnitude, it first must be able to calculate the earthquake's epicentral location. This can only be done if seismographic records from at least three different stations are available. In situations where only one or two records are available the software cannot locate the epicentre. Consequently, in cases where an earthquake cannot be located, determination of its magnitude using EQLOCL has always been problematic.

This appendix describes a method of extracting parametric information from past earthquake magnitudes, located with the SRC software using FS03 seismograms, that can be used in a suitable mathematical formula to determine the magnitude of other earthquakes recorded on the FS03 seismograph, using information from the single station data. This allows the magnitude determination to be done independent of the SRC software.

#### Background Information

The Richter local earthquake magnitude ( $M$ ) is calculated using the formula given in Eq. 1.

$$M = \log_{10}A - \log_{10}A_0 \quad (\text{Eq. 1})$$

Where:

$A$  is the maximum amplitude of the seismic record of the earthquake, and

$A_0$  is the maximum amplitude that would be produced on the same sensor by an earthquake of magnitude zero, occurring at the same location as the earthquake under consideration.

The value of  $\log_{10}A_0$  is dependent only on the epicentral distance of the earthquake from the sensor, and the response characteristics of the sensor itself. It is assumed that the relationship is as given in Eq. 2 (**NOTE: This assumed relationship has yet to be confirmed as being valid**).

$$\log_{10}A_0 = a\delta + b \quad (\text{Eq. 2})$$

Where:

$\delta$  is the epicentral distance, and

a and b are parameters yet to be determined, characteristic of the sensor.

### Method

It is clear that Eq.2 is linear. Therefore the sensor parameters a and b can be determined from the slope and intercept, respectively, of the graph of  $\log_{10}A_0$  plotted against  $\delta$ , providing that sufficient data is available

The epicentral distance  $\delta$  can be expressed in any value that provides a valid determination of the distance from the sensor to the epicentre. This could be (for example):

- the difference in arrival times of the P and S waves (in seconds for example); or,
- the surface distance from sensor to epicentre (in km for example); or,
- the Earth centric angle of arc from sensor to epicentre (in degrees for example).

The values for  $\log_{10}A_0$  can be calculated from past earthquake events, the magnitudes of which have been determined with the SRC software using FS03 seismograms.

Transformation of Eq.1 gives Eq. 3.

$$\log_{10}A_0 = \log_{10}A - M \quad (\text{Eq.3})$$

Table 2 presents the calculations of  $\log_{10}A_0$  based on nine past events that were quantified with the SRC software, showing the S-P time differences used to measure epicentral distances.

**Table 2: Determination of  $\log_{10}A_0$  from past events recorded on the FS03 seismograph.**

Earthquake Date	Measured P arrival in relative seconds	Measured S arrival in relative seconds	S-P time (s)	Measured Amplitude A	Magnitude estimated using ES&S algorithm M	Calculated $\log_{10}(A_0)$
2012-09-19 06:14	11.54	14.97	3.43	1900	1.6	1.6787536
2012-05-20 17:58	42.23	45.79	3.56	1103	1.5	1.5425755
2012-09-22 23:59	38.31	41.91	3.6	243	1.0	1.3856063
2012-04-10 01:51	37.54	42.57	5.03	473.2	1.4	1.2750447
2012-09-25 03:06	10.56	22.7	12.14	456	1.9	0.7589648
2012-08-19 22:37	29.38	41.82	12.44	215	1.5	0.8324385
2012-09-03 15:04	10.86	26.84	15.98	1828	2.8	0.4619762
2012-09-23 16:29	36.21	53.77	17.56	3620	3.2	0.3587086
2012-01-05 14:05	9.96	56.18	46.22	1352	4.3	-1.1690233

Figure 4 shows the graph of  $\log_{10}A_0$  plotted against the associated S-P time difference (extracted from Table 1).




Figure 4:  $\log_{10}A_0$  Vs S-P

Figure 4 also displays the line of best fit, calculated using linear regression of the plotting data, along with the slope, intercept, and correlation coefficient ( $R^2$ ). **The  $R^2$  value of 0.97 confirms that the assumed linear relationship is valid.**

By substituting the slope and intercept values into Eq.1 and Eq.2 we arrive at the formula for FS03 magnitudes given in Eq.4.

$$M_{FS03} = \log_{10}A - (-0.064(S - P) + 1.64) \quad (\text{Eq.4})$$

Where:

$M_{FS03}$  is the Richter magnitude determined from an FS03 seismogram record;

A is the maximum amplitude of the unfiltered FS03 seismogram record;

S is the arrival time of the S wave in seconds, and;

P is the arrival time of the P wave in seconds.

### Important Note Concerning Accuracy and Precision

Table 2, Figure 4, and Equation 4, show a shortened calculation using only 9 historical events, to demonstrate the method. A consequence of using so few input values is that the resulting error ranges will suffer. Consequently, in order to reduce the standard errors in magnitude calculations

based on this method, and extend the accuracy to at least one decimal point, many more input data are required.

The calculations used to determine the actual  $\log_{10}A_0$  values for FS03, used in quantifying earthquake magnitudes, used 34 historical events. This resulted in parameter **a** and **b** values for Equation 2, as shown in Table 3.


**Table 3: Equation 2, a and b Parameter values and Standard Errors.**

	<b>a</b>	<b>b</b>
<b>Estimation</b>	-0.088	1.81
<b>Standard Error</b>	$\pm 0.004$	$\pm 0.05$
<b>Correlation</b>	0.94	

This implies that magnitudes determined using this method will be accurate to at least one decimal place. The a and b values shown in Table 3 are those used at CQSRG to calculate local magnitudes of events recorded by station FS03.

### Example Usage

Figure 2 shows the seismogram of an earthquake that was recorded on station FS03 on 26 October 2012.



**Figure 5: FS03 record of an earthquake.**

From Figure 2 we can obtain the maximum amplitude ( $A = 460$ ), the P wave arrival time ( $P = 25.69$  s) and the S wave arrival time ( $S = 27.90$  s); from which the time difference ( $S - P = 2.21$  s) can be determined.

Inserting these values into Eq.4 we calculate a Richter magnitude of 1.2 (rounded to one decimal place).


Table 2 shows the results of some other similar calculations, for different earthquakes.

**Table 4: Calculations of FS03 Richter magnitudes for some earthquakes.**

Earthquake Date	Measured P arrival in relative seconds	Measured S arrival in relative seconds	S-P time (s)	Measured Amplitude A	Calculated $M_{FS03}$ Magnitude
2012-09-28 16:38	10.56	22.7	12.14	304	1.6
2012-10-03 17:29	25.09	27.69	2.6	259	0.9
2012-10-18 14:48	23.43	25.91	2.48	911	1.5
2012-10-26 04:47	25.69	27.9	2.21	460	1.2

### Student Resources

Figures 6, 7, 8 and 9 are images of earthquake seismograms recorded by FS03. They are included here for the reader to use as practice on the CQSRG magnitude determination method. They can also be used as a resource for High School science teachers who may want to use the formulae presented here as real-world examples of applied mathematics.



**Figure 6: Earthquake recorded on FS03 on 28 September 2012**




Figure 7: Earthquake recorded on FS03 on 3 October 2012




Figure 8: Earthquake recorded on FS03 on 18 October 2012




Figure 9: Earthquake recorded on FS03 on 26 October 2012